



الرقم: مياھنا / ٧٥٨ / ك.ج. التاريخ: ٢١ / ٥ / 2022
Ref.: Date:

السادة :

فاكس :

هاتف :

الموضوع : الملحق رقم (1) للمناقصة رقم C-P-22-013-A

تحية طيبة وبعد،،

بخصوص المناقصة أعلاه والمتعلقة بتوريد وتركيب جهاز Electromagnetic Flow Meter، يرجى اعتماد المواصفات المرفقة مع تعديل اسم المناقصة ليصبح Supply and Installation of Ultrasonic flow meters.

وعليه فقد تقرر تمديد مدة البيع لغاية 2022/06/15 وموعد استلام وفتح العروض يوم الاحد الموافق 2022/06/19 لغاية الساعة الحادية عشر بدلاً من يوم الاحد الموافق 2022/06/05.

مع الإحترام والتقدير

مدير مديرية العطاءات والمشتريات

السيد احمد اللوزي

*المرفقات: مواصفات خاصة صفحتين

Specification Sheet

Supply and Installation of Ultrasonic flow meters

Specifications:

Type	Ultrasonic Clamp-on
Application	Raw water with powder activated carbon (not solvent)
Pipe diameter	900 mm
Flow range	(0 - 120) m3/min
Operational pressure	5 bar
power supply	95 to 240 V_AC 24 V_DC
Inlet run	Must be suitable for limited spaces (short inlet run space)
Keypad	Keypad configurable: A choice of engineering parameters
Diagnostic function	Test mode and self-diagnostics
Current Output	(4-20)mA
Pulse output	forward flow
IP	IP67 for transmitter and I68 for the sensor
Standard measurement accuracy	±(0.5 – 2)%
Liquid temperature range	+0.1°C / +50°C
Ambient temperature	-10...+50°C

Special requirements:

- **The calibration shall be conducted at factory, the calibration certificate and/or calibration report shall be shipped with the flow meter.**
- **The Supplier shall be responsible for all installation requirement**
- **To the site visit arrangement with Eng.Ehab Awwad 0795427657**

- Minimum 1 year warranty period
- Delivery within 12 weeks
- Quantity 1 pieces will be install in Zai water treatment plant
- USA ,Japan or European Brand-name origin
- Offer should include the entire technical specification data sheet.
- All certificates should be submitted along the case of award.